



dexipro[®]X



EN ISO 374-5:2016	LEVEL	100 gloves per box	10 boxes per case
Protection against Bacteria and Fungi	PASS		
Protection against Viruses	PASS		

120 cases per pallet

Single Wall Thickness (mm)	Width (mm)	Length (mm)
0.14±0.03 0.10±0.03 0.09±0.03	- S 80±10 - M 95±10 - L 110±10 - XL 120±10	min. 240

Medical Device Regulation (UK) 2002 (as amended)	CLASS I	Break Force
Medical Device Regulation (EU) 2017/745	CLASS I	Unaged Aged
PPE Regulation (EU) 2016/425	CATEGORY III	Min. 6N Min. 6N
EN ISO 374-1:2016+A1:2018	LEVEL	EN ISO 374-4:2019 Degradation %
40% Sodium Hydroxide (K)	6	-68.1
30% Hydrogen Peroxide (P)	2	30.5
37% Formaldehyde (T)	5	9.5
EN ISO 374-1:2016+A1:2018 Chemical Permeation	1 2 3 4 5 6	
Breakthrough Time (Minutes)	>10 >30 >60 >120 >240 >480	

Regulations	PPE Regulation (EU) 2016/425
	Medical Device Regulation (EU) 2017/745, 93/42/EEC
Quality Standards	ISO 9001:2015 Quality Management Systems
	ISO 13485: 2016 Medical Devices Quality Management Systems
Harmonised Standards	BS EN ISO 374 - 1,2,4,5 Protective gloves against dangerous chemicals and microorganisms
	EN 21420: 2020 Protective gloves. General requirements and test methods
	BS EN 1186: 2002. Materials and articles in contact with foodstuffs
	BS EN 455 1-4 Medical gloves for single use

SMALL	MEDIUM	LARGE	X-LARGE
71996	71997	71998	71999

Legal Manufacturer: SUPERMAX HEALTHCARE LIMITED
 12 - 16 Titan Drive, Fengate, Peterborough, PE1 5XN, United Kingdom
 SUPERMAX HEALTHCARE EUROPE LIMITED
 38 Main Street, Swords, Co. Dublin, Ireland, K67 E0A2

dexipro.co.uk



dexipro[®]X